



Australian Government

Department of Health and Aged Care
Therapeutic Goods Administration

Australian Register of Therapeutic Goods Certificate

Issued to

Soniclean Pty Ltd

for approval to supply

Soniclean Pty Ltd - Test instrument, <specify>

ARTG Identifier	403455
ARTG Start Date	25/01/2023
Product Category	Medical Device Included Class 1
GMDN	36613
GMDN Term	Test instrument, <specify>
Intended Purpose	<p>A range of instruments used to test the performance of Ultrasonic Cleaners:</p> <p>1)UPM: The UPM260 is an Electronic device that provides an indication of the performance of an Ultrasonic cleaner, providing a relative measurement of ultrasonic power present in the liquid within the tank</p> <p>2)SoniClips: These are the reusable clips that helps the users to identify the performance of the Ultrasonic cleaners</p> <p>Both the products are essential tools to track the performance and efficacy of their Ultrasonic Cleaning device, ensuring the confidence that it is operating as promised. This also be replaces the widely used foil test.</p>

Manufacturer Details	Address	Certificate number(s)
Soniclean Pty Ltd	12B Islington Court Dudley Park , SA , 5008 Australia	

ARTG Standard Conditions

The above Medical Device Included Class 1 has been entered on the Register subject to the following conditions:

- The inclusion of the kind of device in the ARTG is subject to compliance with all conditions placed or imposed on the ARTG entry. Refer Part 4-5, Division 2 (Conditions) of the Therapeutic Goods Act 1989 and Part 5, Division 5.2 (Conditions) of the Therapeutic Goods (Medical Devices) Regulations 2002 for relevant information.
- Breaching conditions of the inclusion related to the device of the kind may lead to suspension or cancellation of the ARTG entry; may be a criminal offence; and civil penalties may apply.

Products Covered by This Entry

1. Test instrument, <specify>

Product Specific Conditions

No specific conditions have been recorded against this entry.